



Technical Services Ltd.

Quality Assurance Dept.

**CERTIFICATE OF COMPLIANCE**

We hereby certify that our products (see list below) were manufactured in accordance with our current specifications of materials and production processes and have been inspected to comply with all quality assurance criteria. All temperature monitoring was performed with temperature logger ALMEMO, model 2590 (serial number H09I00824), verified by A.COM calibration laboratory certified ISO/IEC Standard 17025:2005 and NIST traceable (see appendix A).

Product Code	Description	Batch Number	Expiry Date
TP278	Blood Temp 10 °C (BT10)	AR1956-3	12/2024

**Reuben Soncino**  
**Chief Quality Officer Timestrip**

**Chief Quality Officer**

**October 27, 2022**

**Date**

**Appendix A**



**Calibration Certificate**

Certificate Number: 3092908

Date of Issue: 27/04/2022

Name and signature of person carried out the test: Izik. Mishan Lab Man. 

Approved signatory typed name:

Moshe. Appleboum CEO 

Customer Name: TIMESTRIP

Customer Address:

**Identification of item calibrated:**

Description: THERMOMETER  
 Manufac. ALMEMO  
 Model: 25906  
 S/N: H09100824  
 Inventory: TM-03

As Found Conditions: Proper

Date of receipt of item: 27/04/2022

Date of performance of test: 27/04/2022

Environment details: TEMPERATURE 23°C±3°C; HUMIDITY 30% - 70%

Testing Description: The calibration process was performed according to Acom procedures and based on Acom traceability. The calibration values for the unit under test was performed by calculate the indicated value and applied value from the reference standards. The calibration performed according to Acom procedures No. 08-10A to 08-70A

**Testing Recommendation**

Description of specific test or method used:

<u>Model</u>	<u>Serial No.</u>	<u>V.O.C.</u>
TESTO 174H	3660256	01/07/2022
OMEGAETTE HH311	171202193	01/07/2022
FLUKE 2605	5129806	01/07/2022
FLUKE 5500A	6715008	01/07/2022
PMI-108	108-M-1001	01/07/2022
PT100	6517	01/07/2022

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%



Technical Services Ltd.

Quality Assurance Dept.



Certificate Number: 3092908

Page 2 of 4

Test Summary:

Test	Status
1) TEMPERATURE M0 (PT100-2)	Pass
2) TEMPERATURE M1 (PT100-2)	Pass
3) TEMPERATURE M2 (PT100-2)	Pass
4) TEMPERATURE M0 (PT100-1)	Pass
5) TEMPERATURE M1 (PT100-1)	Pass
6) TEMPERATURE M2 (PT100-1)	Pass

**TEMPERATURE M0 (PT100-2)**

Test Point	Applied	Indicated Value	Difference	Tolerance	Uncertainty	Status
-25°C	-25.00	-25.15	-0.15	± 0.75	0.3	Pass
0°C	0.00	-0.17	-0.17	± 0.75	0.3	Pass
25°C	25.00	24.88	-0.12	± 0.75	0.3	Pass
50°C	50.00	49.84	-0.16	± 0.75	0.3	Pass
75°C	75.00	74.55	-0.45	± 1	0.4	Pass
100°C	100.00	99.97	-0.03	± 1	0.4	Pass

**TEMPERATURE M1 (PT100-2)**

Test Point	Applied	Indicated Value	Difference	Tolerance	Uncertainty	Status
-25°C	-25.00	-25.16	-0.16	± 0.75	0.3	Pass
0°C	0.00	-0.11	-0.11	± 0.75	0.3	Pass
25°C	25.00	24.86	-0.14	± 0.75	0.3	Pass
50°C	50.00	49.99	-0.01	± 0.75	0.3	Pass
75°C	75.00	74.93	-0.07	± 1	0.4	Pass
100°C	100.00	99.91	-0.09	± 1	0.4	Pass

**TEMPERATURE M2 (PT100-2)**

Test Point	Applied	Indicated Value	Difference	Tolerance	Uncertainty	Status
-25°C	-25.00	-25.12	-0.12	± 0.75	0.3	Pass
0°C	0.00	-0.01	-0.01	± 0.75	0.3	Pass
25°C	25.00	24.98	-0.02	± 0.75	0.3	Pass
50°C	50.00	49.92	-0.08	± 0.75	0.3	Pass
75°C	75.00	74.81	-0.19	± 1	0.4	Pass
100°C	100.00	99.84	-0.16	± 1	0.4	Pass

**TEMPERATURE M0 (PT100-1)**

Test Point	Applied	Indicated Value	Difference	Tolerance	Uncertainty	Status
-25°C	-25.00	-25.26	-0.26	± 0.75	0.3	Pass
0°C	0.00	-0.13	-0.13	± 0.75	0.3	Pass
25°C	25.00	24.84	-0.16	± 0.75	0.3	Pass
50°C	50.00	49.81	-0.19	± 0.75	0.3	Pass
75°C	75.00	74.83	-0.17	± 1	0.4	Pass
100°C	100.00	99.81	-0.19	± 1	0.4	Pass

**TEMPERATURE M1 (PT100-1)**

Test Point	Applied	Indicated Value	Difference	Tolerance	Uncertainty	Status
-25°C	-25.00	-25.1	-0.10	± 0.75	0.3	Pass
0°C	0.00	-0.14	-0.14	± 0.75	0.3	Pass
25°C	25.00	24.86	-0.14	± 0.75	0.3	Pass
50°C	50.00	49.9	-0.10	± 0.75	0.3	Pass
75°C	75.00	74.87	-0.13	± 1	0.4	Pass
100°C	100.00	99.88	-0.12	± 1	0.4	Pass

**TEMPERATURE M2 (PT100-1)**

Test Point	Applied	Indicated Value	Difference	Tolerance	Uncertainty	Status
-25°C	-25.00	-25.04	-0.04	± 0.75	0.3	Pass
0°C	0.00	-0.15	-0.15	± 0.75	0.3	Pass
25°C	25.00	24.56	-0.44	± 0.75	0.3	Pass
50°C	50.00	49.85	-0.15	± 0.75	0.3	Pass
75°C	75.00	74.86	-0.14	± 1	0.4	Pass
100°C	100.00	99.9	-0.10	± 1	0.4	Pass

A.com certifies that the performance of the above instrument has been verified using test equipment and known accuracy which is traceable to the U.S. National Institute of Standards and Technology (NIST). The policies and procedures at this facility are based on ANSI/ISO/IEC Standard 17025:2017. This certificate shall not be reproduced, except in full without the written approval of the calibration facility.