



Technical Services Ltd.

Quality Assurance Dept.

CERTIFICATE OF COMPLIANCE

We hereby certify that our products (see list below) were manufactured in accordance with our current specifications of materials and production processes and have been inspected to comply with all quality assurance criteria. All temperature monitoring was performed with temperature logger ALMEMO, model 2590 (serial number H09I00824), verified by A.COM calibration laboratory certified ISO/IEC Standard I7025:2005 and NIST traceable (see appendix A).

Product Code	Description	Batch Number	Expiry Date
TP278	Blood Temp 10 °C (BT10)	AR2092-4	11/2028


Reuben Soncino
Chief Quality Officer Timestrip

Chief Quality Officer

June 29, 2026
Date

Appendix A



14 May 2026

Certificate of Calibration

Certificate No./Rev.: 2026-6161/First

Customer Name	Timestrip Technical Services Limited
Customer Address	1 Haazmaut Street, Even Yehuda, 4050269

	Instrument		
Manufacturer	AHLBORN		
Model	ALMEMO 2590-3S	PT100-1	PT100-2
Serial No.	Controller H09100824	Probe 1 ZA 9030-FS1	Probe 2 ZA 9030-FS2
I.D. No.	TM-03	TM-03 PT100-1	TM-03 PT100-2

Item Condition	In Order	Calibration Location	At AI-Pro Laboratory
Received		10.05.2026	
Calibration Date		11-15.05.2026	
Recommendation for Next Calibration Date		05.2027	

	Description	Model	S.N.	Due Date
Reference Measurement Standards Used	PC	N/A	N/A	N/A
Temperature Instrument	IRTD	M2801	F1213	09.2026

Ambient Conditions Temperature 20.0°C ± 2.0°C

Conclusion	Meet the specification. Calibrated. Items were calibrated in loop – instruments are not calibrated separately.
-------------------	---

Performed By: Serge Saada

Approved By: **Itai Amrofel**

Calibrations are in compliance with ISO/IEC 17025:2017. Performed under working instruction WI-IT-01. Calibrated results related only to the calibrated item. This calibration was conducted using standards traceable to the SI through accredited calibration laboratories or national/international metrology institutes. Method performance with Traceability of measurement to the system of units and/or to units of measurement realized at the National Physical Laboratory (UK), NIST or Other recognized national metrology institutes. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of $k=2$. Calibration due dates appearing on the calibration certificate and label are determined by the customer for administrative purposes. Binary statement of conformity where measurement error plus uncertainty is less than MPE (Criteria). The calibration certificate may not be reproduced other than in full and it without signature is not valid. Moreover, the calibration for vehicles, chambers or equipment performed with compliance to SI1291 Israeli Standard Sections 3.3.4 and 3.4.1.



Calibration:

AS Found:

Temperature Results For Probe 1 TM-03 PT100-1 (S.N. ZA 9030-FS1):

Nominal	Reference Value °C	Measurement Results °C	Measurement Error °C	Expanded Uncertainty °C	Criteria	Status
-25.0	-24.974	-24.7	0.3	0.2	0.5	Pass
0.0	0.016	0.1	0.1	0.2	0.5	Pass
25.0	24.999	25.1	0.1	0.2	0.5	Pass
50.0	49.990	50.2	0.2	0.2	0.5	Pass
100.0	100.006	99.8	0.2	0.2	0.5	Pass

Temperature Results For Probe 2 TM-03 PT100-2 (S.N. ZA 9030-FS2):

Nominal	Reference Value °C	Measurement Results °C	Measurement Error °C	Expanded Uncertainty °C	Criteria	Status
-25.0	-24.974	-25.33	0.36	0.1	0.5	Pass
0.0	0.016	-0.31	0.33	0.1	0.5	Pass
25.0	24.999	24.79	0.21	0.1	0.5	Pass
50.0	49.990	49.90	0.09	0.1	0.5	Pass
100.0	100.006	100.18	0.12	0.1	0.5	Pass

AS Left:

Parameters For Probe 1 (Channel 0-2) TM-03 PT100-1 (S.N. ZA 9030-FS1):

Zero correct	Slope correct
0.6°C	1.0000°C

Temperature Results For Probe 1 TM-03 PT100-1 (S.N. ZA 9030-FS1):

Nominal	Reference Value °C	Measurement Results °C	Measurement Error °C	Expanded Uncertainty °C	Criteria	Status
-25.0	-24.927	-24.8	0.13	0.2	0.5	Pass
0.0	0.076	0.2	0.12	0.2	0.5	Pass
25.0	25.089	25.2	0.11	0.2	0.5	Pass
50.0	50.078	50.2	0.12	0.2	0.5	Pass
100.0	100.028	100.1	0.07	0.2	0.5	Pass

Parameters For Probe 2 (Channel 0-2) TM-03 PT100-2 (S.N. ZA 9030-FS2):

Zero correct	Slope correct
0.09°C	1.0027°C

Temperature Results For Probe 2 TM-03 PT100-2 (S.N. ZA 9030-FS2):

Nominal	Reference Value °C	Measurement Results °C	Measurement Error °C	Expanded Uncertainty °C	Criteria	Status
-25.0	-24.927	-24.83	0.12	0.1	0.3	Pass
0.0	0.076	0.06	0.02	0.1	0.3	Pass
25.0	25.089	25.05	0.04	0.1	0.3	Pass
50.0	50.078	50.07	0.01	0.1	0.3	Pass
100.0	100.028	100.03	0.00	0.1	0.3	Pass

Note: The Probes were tested at each of the three channels and the results are the same (the probes are connected with a transmitter to the controller).